

Specifications

Weights and Electrical Specifications

Patient Weight (maximum):	450 lbs (204.1 kg)
Weight of Chair: 641 (w/standard base) w/packaging & skid	440 lbs (200 kg) 495 lbs (224 kg)
641 (w/rotational base) w/packaging & skid	525 lbs (238 kg) 585 lbs (265 kg)
Power Cord:	8 ft. (243.8 cm) long
Electrical Requirements:	[See Model I.D. / Compliance Charts]
Foot / Hand Control Voltage:	10 VAC, SELV (Safety Extra Low Voltage)
Duplex Receptacle (optional) maximum load:	115 VAC, 3A, 50/60 Hz
Fuses: <i>Located at power cord inlet: F1 & F2</i>	6.3 A, 250V, Type T, 5 x 20 mm
Duty Cycle (motor run time):	Intermittent Operation [30 seconds ON / 5 minutes OFF]
Classifications:	Class 1, Type B Applied Part
Protection against ingress of fluids: <i>Foot Control <u>only</u>:</i>	Ordinary Equipment IPX1
Regulatory Compliance:	[See Model I.D. / Compliance Charts]



WARNING

Equipment is not suitable for use in the presence of a flammable anesthetic mixture.

Clarification: Equipment is suitable for use in the presence of oxygen, air, or nitrous oxide.